



Handläggande organ

Standardiseringsgruppen STG

Fastställt

1999-09-10

Utgåva

1

Sida

1 (1+10)

© Copyright SIS. Reproduction in any form without permission is prohibited.

Laboratory glassware – One-mark volumetric flasks (ISO 1042:1998)

The European Standard EN ISO 1042:1999 has the status of a Swedish Standard. This document contains the official English version of EN ISO 1042:1999.

Swedish Standards corresponding to documents referred to in this Standard are listed in "Catalogue of Swedish Standards", issued by SIS. The Catalogue lists, with reference number and year of Swedish approval, International and European Standards approved as Swedish Standards as well as other Swedish Standards.

Laboratorieglass – Mätkolvar (ISO 1042:1998)

Europastandarden EN ISO 1042:1999 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 1042:1999.

Motsvarigheten och aktualiteten i svensk standard till de publikationer som omnämns i denna standard framgår av "Katalog över svensk standard", som ges ut av SIS. I katalogen redovisas internationella och europeiska standarder som fastställts som svenska standarder och övriga gällande svenska standarder.

ICS 17.060; 71.040.20

Standarder kan beställas hos SIS som även lämnar allmänna upplysningar om svensk och utländsk standard.
Postadress: SIS, Box 6455, 113 82 STOCKHOLM
Telefon: 08 - 610 30 00. Telefax: 08 - 30 77 57

Upplysningar om **sakinnehållet** i standarden lämnas av STG.
Telefon: 08 - 13 62 50. Telefax: 08 - 618 61 28
E-post: info@stg.se
Prisgrupp N

Tryckt i oktober 1999

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 1042

May 1999

ICS 17.060

English version

**Laboratory glassware - One-mark volumetric flasks (ISO
1042:1997)**

Verrerie de laboratoire - Fioles jaugées à un trait (ISO
1042:1997)

Labogeräte aus Glas - Meßkolben (ISO 1042:1997)

This European Standard was approved by CEN on 4 April 1999.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Foreword

The text of the International Standard from Technical Committee ISO/TC 48 "Laboratory glassware and related apparatus" of the International Organization for Standardization (ISO) has been taken over as an European Standard by Technical Committee CEN/TC 332 "Laboratory equipment", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 1999, and conflicting national standards shall be withdrawn at the latest by November 1999.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of the International Standard ISO 1042:1998 has been approved by CEN as a European Standard without any modification.

Introduction

Volumetric flasks together with analytical balances are the fundamental tools for the preparation of volumetric standard solutions – the basis of chemical analysis. The design of narrow-necked class A volumetric flasks has been optimized to achieve the fewest possible acceptable errors.

With the increasing popularity of piston-operated pipettors, there is market pressure for the manufacture of volumetric flasks with wider necks so that pipettor tips may be inserted to remove solution directly. Wide-necked flasks will of necessity be of lower accuracy than the corresponding capacities of narrow-necked flasks and the insertion of any extraneous device may introduce other errors.

It is therefore recommended that narrow-necked class A volumetric flasks are used for the production of standard solutions and where necessary, a suitable quantity should be poured into an intermediate vessel into which the pipettor tip may be introduced.

In accordance with good laboratory practice, only narrow-necked class A volumetric flasks conforming to this International Standard should be used for precise analytical purposes.

Laboratory glassware — One-mark volumetric flasks

1 Scope

This International Standard specifies requirements for an internationally acceptable series of one-mark volumetric flasks, suitable for general laboratory purposes.

The specifications in this International Standard are in conformity with ISO 384 and with OIML Recommendation No. 4.

2 Normative references

The following standards contain provisions which, through references in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 383:1976, *Laboratory glassware — Interchangeable conical ground joints.*

ISO 719:1985, *Glass — Hydrolytic resistance of glass grains at 98 °C — Method of test and classification.*

ISO 4787:1984, *Laboratory glassware — Volumetric glassware — Methods for use and testing of capacity.*

3 Basis of adjustment

3.1 Unit of volume

The unit of volume shall be the millilitre (ml) which is equivalent to the cubic centimetre (cm³).

NOTE — The term millilitre (ml) is commonly used as a special name for the cubic centimetre (cm³), in accordance with a decision of the twelfth Conférence Générale des Poids et Mesures. The term millilitre is acceptable, in general, for references in International Standards to capacities of volumetric glassware and it is used, in particular, in the present text.

3.2 Reference temperature

The standard reference temperature, i.e. the temperature at which the volumetric flask is intended to contain its nominal volume (nominal capacity), shall be 20 °C.

When the flask is required for use in a country which has adopted a standard reference temperature of 27 °C, however, this value shall be substituted for 20 °C.