

Laboratorieglass – Provrör
(ISO 4142:2002)

Laboratory glassware – Test tubes
(ISO 4142:2002)

Europastandarden EN ISO 4142:2002 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 4142:2002.

The European Standard EN ISO 4142:2002 has the status of a Swedish Standard. This document contains the official English version of EN ISO 4142:2002.

Dokumentet består av 8 sidor.

Upplysningar om **sakinnehållet** i standarden lämnas av SIS, Swedish Standards Institute, tel 08 - 555 520 00.

Standarder kan beställas hos SIS Förlag AB som även lämnar **allmänna upplysningar** om svensk och utländsk standard.

Postadress: SIS Förlag AB, 118 80 STOCKHOLM
Telefon: 08 - 555 523 10. *Telefax:* 08 - 555 523 11
E-post: sis.sales@sis.se. *Internet:* www.sis.se

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 4142

August 2002

ICS 71.040.20

English version

Laboratory glassware - Test tubes (ISO 4142:2002)

Verrerie de laboratoire - Tubes à essais (ISO 4142:2002)

Laborgeräte aus Glas - Reagenzgläser (ISO 4142:2002)

This European Standard was approved by CEN on 10 July 2002.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

This document (EN ISO 4142:2002) has been prepared by Technical Committee ISO /TC 48 "Laboratory glassware and related apparatus" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2003, and conflicting national standards shall be withdrawn at the latest by February 2003.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 4142:2002 has been approved by CEN as EN ISO 4142:2002 without any modifications.

Laboratory glassware — Test tubes

1 Scope

This International Standard specifies a range of test tubes, suitable for general laboratory use, fabricated from borosilicate, neutral or soda/lime glass, which are designated Type I, Type II and Type III respectively.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 695, *Glass — Resistance to attack by a boiling aqueous solution of mixed alkali — Method of test and classification*

ISO 719, *Glass — Hydrolytic resistance of glass grains at 98 °C — Method of test and classification*

ISO 1776:1985, *Glass — Resistance to attack by hydrochloric acid at 100 °C — Flame emission or flame atomic absorption spectrometric method*

ISO 3585, *Borosilicate glass 3.3 — Properties*

ISO 4803, *Laboratory glassware — Borosilicate glass tubing*

3 Classification and designation

The following three types of test tubes are specified.

- Type I test tubes (borosilicate glass) are suitable for most usual laboratory applications. They will stand most temperatures commonly used, including boiling of samples. These test tubes are highly chemically resistant.
- Type II test tubes (neutral glass) are suitable for many less demanding applications, and will withstand moderate warming, e. g. in a water bath, and/or moderate temperature change. They should not be taken from the cold and placed directly into the hottest part of the flame without being preheated. Type II test tubes are chemically resistant and are suitable for use with samples which are susceptible to pH changes.
- Type III test tubes (soda/lime glass) are suitable for general mixing and simple laboratory work, and will withstand moderate warming, e. g. in a water bath, and/or moderate temperature change. They should not be placed into naked flames. Their chemical resistance is limited.

These type numbers are specific to this International Standard and should not be confused with similar numbering used in the hydrolytic resistance classification.

If a designation of test tubes is required, this shall be by reference to this International Standard ISO 4142, together with the type designation, the nominal size and the wall thickness of the test tube.

EXAMPLE For a test tube Type I with a nominal size of 10 mm × 75 mm and a medium wall thickness of 1,0 mm, the designation would be as follows:

Test tube ISO 4142-10×75-M

4 Material

4.1 Test tubes shall be of clear glass as free as possible from visible defects and internal stress.

4.2 Type I test tubes shall be manufactured from borosilicate glass in accordance with ISO 3585 and ISO 4803, with coefficient of thermal expansion of $3,3 \times 10^{-6} \text{K}^{-1}$.

4.3 Type II test tubes shall be manufactured from neutral glass as specified by the manufacturer. Typically the coefficient of thermal expansion will be $5,0 \times 10^{-6} \text{K}^{-1}$.

4.4 Type III test tubes shall be manufactured from soda/lime glass as specified by the manufacturer. Typically the coefficient of thermal expansion will be $9,1 \times 10^{-6} \text{K}^{-1}$.

4.5 Water resistance shall be in accordance with ISO 719, alkali resistance in accordance with ISO 695 and acid resistance in accordance with ISO 1776. The resistance classes of the glass types shall comply with Table 1.

Table 1 — Minimum requirements for chemical resistance

Property	Minimum requirement		
	Type I	Type II	Type III
Water resistance	Class HGB 1	Class HGB 1	Class HGB 3
Acid resistance	$\leq 100 \mu\text{g Na}_2\text{O}^{\text{a}}$	$\leq 100 \mu\text{g Na}_2\text{O}^{\text{a}}$	—
Alkali resistance	Class A2	Class A2	Class A2

^a See clause 9 in ISO 1776:1985.

5 Construction and dimensions

5.1 The top (open end) of each test tube shall be smoothly finished at right angles to the axis, with either a rim or a fire-polished end.

5.2 Dimensions shall be in accordance with the lengths, diameters and tolerances shown in the Table 2 to Table 4.

5.3 The bottom of each test tube shall be essentially hemispherical, with a wall thickness of no less than 67 % and no more than 167 % of the nominal side wall thickness.

6 Marking

6.1 Each Type I test tube shall be marked to clearly indicate that it is manufactured from borosilicate glass with a coefficient of thermal expansion of $3,3 \times 10^{-6} \text{K}^{-1}$, e.g. "boro 3.3". The manufacturer's or vendor's name or mark may also be applied to each test tube.

6.2 In the case of Type II and Type III test tubes, the appropriate information may be on the packaging only.