

SVENSK STANDARD

SS-ISO 22241-2:2019

Fastställt/Approved: 2019-02-21
Utgåva/Edition: 2
Språk/Language: engelska/English
ICS: 43.060.40

Dieselmotorer – Tillsats AUS 32 för reduktion av NOx – Del 2: Provningsmetoder (ISO 22241-2:2019, IDT)

Diesel engines – NOx reduction agent AUS 32 – Part 2: Test methods (ISO 22241-2:2019, IDT)

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-80010139>

Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Den internationella standarden ISO 22241-2:2019 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av ISO 22241-2:2019.

Denna standard ersätter SS-ISO 22241-2:2006, utgåva 1 och SS-ISO 22241-2:2006/Cor 1:2008, utgåva 1

The International Standard ISO 22241-2:2019 has the status of a Swedish Standard. This document contains the official English version of ISO 22241-2:2019.

This standard supersedes the SS-ISO 22241-2:2006, edition 1 and SS-ISO 22241-2:2006/Cor 1:2008, edition 1

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS som även lämnar allmänna upplysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS, who can also provide general information about Swedish and foreign standards.

Denna standard är framtagen av kommittén för Motorer i bilar, SIS/TK 220.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

Contents

Page

Foreword	v
1 Scope.....	6
2 Normative references	6
3 Terms and definitions.....	6
4 Specifications.....	6
5 Sampling.....	7
6 Precision and dispute	7
Annex A (normative) Sampling.....	8
Annex B (normative) Determination of urea content by total nitrogen	10
Annex C (normative) Refractive index and determination of urea content by refractive index.....	14
Annex D (normative) Determination of alkalinity	17
Annex E (normative) Determination of biuret content	20
Annex F (normative) Determination of aldehyde content.....	25
Annex G (normative) Determination of insoluble matter content by gravimetric method	29
Annex H (normative) Determination of phosphate content by photometric method.....	32
Annex I (normative) Determination of trace element content (Al, Ca, Cr, Cu, Fe, K, Mg, Na, Ni, P and Zn) by ICP-OES method.....	38
Annex J (informative) Determination of identity by FTIR spectrometry method	46
Annex K (informative) Precision of test method.....	48
Bibliography	49

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 22, *Road vehicles*, Subcommittee SC 34, *Propulsion, powertrain and powertrain fluids*.

This second edition cancels and replaces the first edition (ISO 22241-2:2006), which has been technically revised. It also incorporates the Technical Corrigendum ISO 22241-2:2006/Cor. 1:2008. The main changes compared to the previous edition are as follows:

- Major revisions to test methods of [Annex C](#) and [Annex I](#),
- Precision values for all test methods were revised,
- [Annex K](#) was updated.

A list of all parts in the ISO 22241 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Diesel engines — NO_x reduction agent AUS 32 —

Part 2: Test methods

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety issues associated with its use. It is the responsibility of users of this document to respond appropriately to ensure the safety and health of personnel prior to application of the document.

1 Scope

This document specifies test methods required for determination of the quality and chemical characteristics of NO_x reduction agent AUS 32 (aqueous urea solution) as specified in ISO 22241-1.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 4259 (all parts), *Petroleum and related products — Precision of measurement methods and results*

ISO 12185, *Crude petroleum and petroleum products — Determination of density — Oscillating U-tube method*

ISO 17034, *General requirements for the competence of reference material producers*

ISO 22241-1, *Diesel engines — NO_x reduction agent AUS 32 — Part 1: Quality requirements*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 22241-1 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 certified reference material CRM

substance or material used to check the quality and metrological traceability of products, to validate analytical measurement methods or for the calibration of instruments

4 Specifications

Conformance with the limits specified in ISO 22241-1 shall be determined by the test methods specified in [Annexes B](#) through [I](#) of this document. If necessary, the identity of the product can be determined as specified in [Annex J](#).

5 Sampling

Samples shall be taken in accordance with [Annex A](#).

6 Precision and dispute

Each of the test methods specified in [Annex B](#) through [Annex I](#) include a precision statement using guidance from ISO 4259 (all parts). In cases of dispute, the procedures described in ISO 4259 (all parts) shall be used for resolving the dispute, and interpretation of the results based on the test method precision shall be used.

For the convenience of the user, the respective precision data are summarized in Table K.1.

Annex A (normative)

Sampling

A.1 General

The sampling method specified in this annex is valid for each sampling of AUS 32 throughout the supply chain after the shipment from the manufacturer's site to the AUS 32 containers of the vehicles.

A.2 Principle

The limits for the quality characteristics of AUS 32, which are specified in ISO 22241-1, are the representative analytical results that can only be obtained when the sample is protected from any contamination before the analysis.

Therefore, suitable bottles shall be used for sampling, which do not contaminate the sample, especially regarding the trace elements.

NOTE The sampling method specified in this annex is based on ISO 5667-3.

A.3 Possible contaminants

During the sampling process, foreign matter can lead to contamination of the sample. Under realistic conditions, the following sources of contamination will pose a major hazard:

- residues of process aids used for the production of the sampling bottles;
- contaminants which have been deposited in the empty bottles during the time they are stored empty;
- contaminants from the air, i.e. dust or any foreign matter from the surrounding, during the sampling;
- residues of cleaning agents, which have been used for cleaning the sampling equipment and the bottles as well;
- fuel.

A.4 Apparatus

A.4.1 Sampling bottles.

1 000-ml wide neck bottles shall be used. Suited materials for these bottles are high density polyethylene (HDPE), high density polypropylene (HDPP), polyvinylidene fluoride (PVDF) and perfluoroalkoxy alkane (PFA).

Prior to the first use with AUS 32, the bottles shall be cleaned and finally rinsed with deionized water followed by AUS 32.

A.4.2 Labels.

Each bottle shall be labelled using labels of approximately 10 cm by 5 cm. The labels and the writing on these labels shall be resistant to water and to AUS 32.

A.5 Sampling

The locked wide-neck bottle shall be opened and the cap shall be placed on a clean surface with the opening turned downward. After flushing the sampling pipe, the bottle shall be filled completely with AUS 32 from the container. The first filling shall be discarded, and the bottle shall immediately be re-filled with AUS 32 and closed tightly. The label shall be attached to the bottle (see [A.4.2](#)). During the filling of the sample, maximum care shall be taken that neither dust nor liquid pollutants get into the bottle.

The filled bottle should reach the laboratory as soon as possible. During transportation and storage, the sample should be kept at the lowest possible temperature, preferably between 0 °C and 15 °C.

It is recommended to conduct the analysis within three weeks in order to take into account possible changes in the ammonia content.

A.6 Sample quantity

The minimum quantity of sample material depends on the type of analysis conducted. Whenever possible, make sure that a sufficient volume of sample material is available (recommendation: 1 litre), and at least double that which is required for complete verification of AUS 32 specifications. In case of dispute, a sufficient number of samples shall be taken according to ISO 4259 (all parts).

A.7 Labelling

The label should contain the following information:

- product name;
- name of the company which owns the sample product¹⁾;
- address where the sample was taken from¹⁾;
- manufacturer of the sample product¹⁾;
- batch or lot number;
- container from which the sample was taken¹⁾;
- part of the container where the sample was taken from (sampling point)¹⁾;
- date and time of sampling¹⁾;
- sample shipment date¹⁾;
- name and signature of the person who took the sample¹⁾.

1) Mandatory only in cases of dispute.

Annex B **(normative)**

Determination of urea content by total nitrogen

B.1 General

This annex specifies the procedure for determining the urea content of AUS 32.

The method is applicable for the determination of the urea content in the range of 30 to 35 % (mass fraction).

B.2 Principle

The sample is combusted at high temperatures in a stream of oxygen. Following the reduction of formed nitrogen oxides to elemental nitrogen and removal of any interfering products of combustion, nitrogen is measured with a thermal-conductivity detector. The urea content is calculated from the determined total nitrogen minus the nitrogen content of biuret.

B.3 Apparatus

B.3.1 Automatic nitrogen analyser, based on combustion methods.

B.3.2 Analytical balance.

The accuracy of the balance is a function of the analyser used and the required weighed portions. Resolution should be 0,1 % of the weighed portion or better.

B.3.3 Auxiliary devices for sample preparation, for example:

- tweezers with a blunt tip;
- micro-spatula with a flattened tip;
- pipette.

The pipette is recommended for weighing in and thus does not need to be calibrated. It is important, however, to obtain a good droplet size (small droplets). Fixed-volume pipettes or pipettes with an adjustable volume in the range from 10 µl to 1 000 µl or single-trip Pasteur pipettes with a fine tip may also be used.

B.3.4 Customary chemically resistant glass.

B.4 Chemicals

B.4.1 Distilled or deionized water, conductivity less than 0,1 mS/m, according to ISO 3696 grade 2.

B.4.2 Auxiliary combustion agent and other equipment, appropriate for use with the selected nitrogen analyser.

The following materials are merely examples. Other or similar materials may be used as required, depending on the system that is available:

- tin capsule or similar sample containers;
- auxiliary combustion agent, non-nitrogenous, such as saccharose, cellulose;
- absorbing agent for liquids, non-nitrogenous, such as magnesium oxide.

B.4.3 Standard substances for nitrogen determination, preferably with certified nitrogen content.

EXAMPLE Suitable standard substances include ethylenediamine tetraacetic acid (EDTA), nicotinic acid amide.

Low-biuret urea of adequate purity (for example crystalline ultra pure or analytical) or other such standard substances recommended by and available from the equipment manufacturer may also be used. Certified standard substances should be preferred.

NOTE Liquid standard substances (e.g. urea solutions) are not suited for calibration purposes.

B.4.4 Oxygen, min. 99,995 % O₂.

B.4.5 Other ultrapure gases, if required to operate the nitrogen analyser, such as helium, min. 99,996 % He.

B.4.6 Other reagents or auxiliary agents, as required by the equipment.

B.5 Procedure

B.5.1 General

The sample should be fully dissolved and free from urea crystals. It may be heated to max. 40 °C as required prior to further processing.

NOTE Different types of apparatus are available on the market. The resulting various resources and modes of operation are not an object of this document. Rather, operation is based on the respective operation manuals.

B.5.2 Reference curve

Perform calibration as required for the specific type of analyser and according to the respective operation manuals (for example, after replacement of the combustion tube, reagent or similar) by performing measurements as described in [B.5.4](#). Weigh in an appropriate amount of standard substances repeatedly as appropriate for the respective types of apparatus to obtain a reference curve.

B.5.3 Inspecting the apparatus for good working order and the reference curve

Use an appropriate standard substance to review the good working order of the apparatus and the reference curve. Preferably, a certified urea standard solution should be used.

Frequency of inspection is a function of the analyser used.

B.5.4 Measurement

Weigh a portion of the sample in a suitable holder (such as a tin capsule) as specified for the type of nitrogen analyser used. The amount should be such that the absolute amount of nitrogen is in the middle range of the reference curve.